


Lighting Passport™

CALIBRATION CERTIFICATE

Model	ALP-01	
Serial No.	1955140010523	
Calibration Date	2014.07.07	
Ambient Temperature	(23±1)°C	
Relative Humidity	(45±10)%	

Item	Standard Value	Check	Δ
Illuminance (lux)	976	977	0.1%
Chromaticity Coordinates	x	0.4373	0.0000
	y	0.4038	0.0000
Specification	$\Delta x, y < \pm 0.002, \Delta \text{lux} < \pm 3\%$		

Calibration Descriptions

1. Location of Test

- This test was performed at 9F., No.43, Fuxing Rd., Xindian Dist., New Taipei City 23150, Taiwan (R.O.C.), Asensetek Lab.

2. Calibration System includes

- Low pressure line lamp for wavelength calibration
- Lighting Source System with NIST traceable driver(#0501132531 LPS-100)
- Spectrophotometer for photometric calibration that traces NIST traceable calibration source (#07099996 IES1000)

Calibration Interval

If the system is handled with care, recalibration is recommended to be made once a year, at the discretion of the user.

QC: Lilia Supervisor: Joe Chen



Address : 8F., No.43, Fuxing Rd., Xindian Dist., New Taipei City
23150, Taiwan (R.O.C.)
Phone : +886-2-2218-5577
FAX : +886-2-2218-0656
E-mail : service@asensetek.com